

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

### AMENDMENTS TO THE CLAIMS

This listing of claims will replace all prior versions, and listings, of claims in the application.

#### Listing of Claims

1. (Currently Amended) A trans-fill system comprising:  
an intensifier operable to increase pressure of therapeutic gas provided from an oxygen concentrator or other source of therapeutic gas, the intensifier creates at an inlet of the intensifier to create a compressed therapeutic gas stream;  
a conserver coupled to the regulated-compressed therapeutic gas stream, the conserver operable to deliver a bolus of therapeutic gas during inhalation of a patient; and  
a cylinder connector operable to couple a portable cylinder to the compressed therapeutic gas stream;  
wherein the trans-fill system is operable to provide therapeutic gas to the cylinder connector to fill the portable cylinder while providing therapeutic gas to the patient through the conserver.
2. (Currently Amended) The trans-fill system as defined in claim 1 further comprising:  
a gas sense device fluidly coupled to the cylinder connector and receiving a portion of the therapeutic gas of the compressed therapeutic gas stream, the gas sense device operable to detect purity of therapeutic gas; and  
wherein the trans-fill system is operable to allow a portion of the gas within a connected portable cylinder to flow to the gas sense device, and wherein the trans-fill system refrains from filling the connected portable cylinder if the purity of the gas in the bottle, as determined by the gas sense device, falls below a predetermined threshold.
3. (Original) The trans-fill system as defined in claim 2 wherein the gas sense device further comprises an oxygen-specific sensor.
4. (Original) The trans-fill system as defined in claim 2 wherein the gas sense device further comprises a density sensor.

Page 4 of 22

135923.01/2230.00400

BEST AVAILABLE COPY

10/04/2004 14:36 FAX 7132388008

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

5. (Original) The trans-fill system as defined in claim 4 wherein the gas sense device further comprises an oxygen-specific sensor.

6. (Currently Amended) A trans-fill system comprising:  
an intensifier operable to increase pressure of therapeutic gas provided at an inlet of the  
intensifier to create a compressed therapeutic gas stream;  
a conserver coupled to the compressed therapeutic gas stream, the conserver operable to  
deliver a bolus of therapeutic gas during inhalation of a patient; and  
a cylinder connector operable to couple a portable cylinder to the compressed therapeutic  
gas stream;  
a gas sense device fluidly coupled to the cylinder connector, the gas sense device  
operable to detect purity of therapeutic gas; and  
~~The trans-fill system as defined in claim 2 further comprising:~~  
a compressor coupled to the cylinder connector, the compressor operable to evacuate  
contents of the connected portable cylinder;  
wherein the trans-fill system is operable to provide therapeutic gas to the cylinder  
connector to fill the portable cylinder while providing therapeutic gas to the  
patient through the conserver;  
wherein the trans-fill system is operable to allow a portion of the gas within a connected  
portable cylinder to flow to the gas sense device, and wherein the trans-fill system  
refrains from filling the connected portable cylinder if the purity of the gas in the  
bottle, as determined by the gas sense device, falls below a predetermined  
threshold; and  
wherein the compressor evacuates the connected portable cylinder if the purity detected  
by the gas sense device falls below the predetermined threshold.

7. (Original) The trans-fill system as defined in claim 6 further comprising:  
an oxygen concentrator providing therapeutic gas to the intensifier; and  
wherein the compressor is part of the oxygen concentrator.

Page 5 of 22

135923.01/2230.00400

PAGE 8/25 \* RCVD AT 10/4/2004 3:32:44 PM [Eastern Daylight Time] \* SVR:USPTO-EFAXRF-1/11 \* DNIS:8729306 \* CSID:7132388008 \* DURATION (mm-ss):07-52

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

8. (Original) The trans-fill system as defined in claim 6 wherein the connected portable cylinder is filled after evacuation.
9. (Currently Amended) A trans-fill system comprising:  
an intensifier operable to increase pressure of therapeutic gas provided at an inlet of the  
intensifier to create a compressed therapeutic gas stream;  
a conserver coupled to the compressed therapeutic gas stream, the conserver operable to  
deliver a bolus of therapeutic gas during inhalation of a patient; and  
a cylinder connector operable to couple a portable cylinder to the compressed therapeutic  
gas stream;  
wherein the trans-fill system is operable to provide therapeutic gas to the cylinder  
connector to fill the portable cylinder while providing therapeutic gas to the  
patient through the conserver; ~~The trans-fill system as defined in claim 1 further~~  
comprising:  
a flow meter coupled on a regulated side to a patient port, and wherein when operating  
the flow meter provides a continuous flow of therapeutic gas to the patient by way  
of the patient port;  
a valve coupled between a source of therapeutic gas and the intensifier, the valve  
selectively couples the source of therapeutic gas to only one of the flow meter and  
the intensifier; and  
wherein the trans-fill system provides therapeutic gas to a patient by one of a continuous  
flow through the flow meter and the bolus delivered by the conserver.
10. (Original) The trans-fill system as defined in claim 9 further comprising:  
a sense circuit mechanically coupled to the flow meter, the sense circuit operable to  
detect a flow setting of the flow meter; and  
a controller electrically coupled to the sense circuit and the conserver;  
wherein the controller is operable to set a bolus volume for delivery to the patient from  
the conserver based on the flow setting of the flow meter.

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

11. (Original) The trans-fill system as defined in claim 1 further comprising:  
a cylinder coupled to the compressed therapeutic gas stream;  
wherein the trans-fill system is operable to provide therapeutic gas from the cylinder  
when the intensifier is not in operation.
12. (Original) The trans-fill system as defined in claim 11 wherein the cylinder is a cylinder  
internal to the trans-fill system.
13. (Original) The trans-fill system as defined in claim 11 wherein the cylinder is external to  
the trans-fill system.
14. (Original) The trans-fill system as defined in claim 13 wherein the cylinder is a portable  
coupled to the cylinder connector.
15. (Currently Amended) A method comprising:  
compressing a stream of low-pressure therapeutic gas to form a compressed therapeutic  
gas stream of sufficient pressure to fill a cylinder;  
providing a first portion of the compressed therapeutic gas stream to fill a cylinder; and  
providing a second portion of the compressed therapeutic gas stream to a patient in-as a  
bolus of therapeutic gas.
16. (Original) The method as defined in claim 15 further comprising, prior to compressing  
step and the providing steps, testing gas within the cylinder.
17. (Original) The method as defined in claim 16 wherein testing gas within the cylinder  
further comprises measuring the oxygen content of gas within the cylinder.
18. (Original) The method as defined in claim 16 wherein testing gas within the cylinder  
further comprises measuring the density of the gas within the cylinder.

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

19. (Original) The method as defined in claim 18 wherein testing the gas within the cylinder further comprises measuring the oxygen content of the gas within the cylinder.
20. (Original) The method as defined in claim 16 further comprising refraining from filling the cylinder if the gas within the cylinder contains contaminants.
21. (Currently Amended) A method comprising:  
testing gas within the cylinder; The method as defined in claim 16 further comprising  
evacuating the contents of the cylinder prior to filling if the gas within the cylinder  
contains contaminants;  
compressing a stream of low-pressure therapeutic gas to form a compressed therapeutic  
gas stream;  
providing a first portion of the compressed therapeutic gas stream to fill a cylinder; and  
providing a second portion of the compressed therapeutic gas stream to a patient as a  
bolus of therapeutic gas.
22. (Original) The method as defined in claim 16 further comprising setting a volume of the bolus of therapeutic gas based on a sensed setting for a continuous flow of therapeutic gas through a flow meter.
23. (Original) A system comprising:  
a compressor operable to increase pressure of therapeutic gas provided at an inlet of the compressor to create a compressed therapeutic gas stream;  
a cylinder coupled to the compressed therapeutic gas stream; and  
a conserver coupled to the compressed therapeutic gas stream, the conserver operable to deliver a bolus of therapeutic gas during inhalation of a patient;  
wherein the system is operable to provide therapeutic gas from the cylinder when the compressor is not in operation.

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

24. (Original) The system as defined in claim 23 further comprising an oxygen concentrator coupled to the compressor, the oxygen concentrator operable to provide therapeutic gas to the compressor.
25. (Original) The system as defined in claim 24 wherein the oxygen concentrator further comprises a pressure swing absorption system.
26. (Original) The system as defined in claim 23 further comprising a connector operable to selectively couple the cylinder to the system.
27. (Original) The system as defined in claim 26 wherein the cylinder is a portable cylinder.
28. (Original) The system as defined in claim 23 wherein the cylinder is internal to the system.
29. (Original) The system as defined in claim 28 further comprising a portable cylinder coupled to the compressed therapeutic gas stream, and wherein the system is operable to provide therapeutic gas from the cylinder and the portable cylinder when the compressor is not in operation.
30. (Original) The system as defined in claim 23 wherein the compressor further comprises an intensifier.
31. (Original) An apparatus comprising:  
an intensifier operable to take therapeutic gas at a first pressure and increase the pressure of the therapeutic gas to a second pressure, higher than the first pressure;  
a fill port fluidly coupled to the therapeutic gas at the second pressure, the fill port operable to selectively couple a cylinder to be filled with therapeutic gas;  
a gas sense device coupled to the fill port, the gas sense device operable to detect content of gas within the cylinder prior to filling.

Page 9 of 22

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

32. (Original) The apparatus as defined in claim 31 wherein the gas sense device further comprises an oxygen-specific sensor.
33. (Original) The apparatus as defined in claim 31 wherein the gas sense device further comprises a density sensor.
34. (Original) The apparatus as defined in claim 33 wherein the gas sense device further comprises an oxygen-specific sensor.
35. (Currently Amended) An apparatus comprising:  
an intensifier operable to take therapeutic gas at a first pressure and increase the pressure  
of the therapeutic gas to a second pressure, higher than the first pressure;  
a fill port fluidly coupled to the therapeutic gas at the second pressure, the fill port  
operable to selectively couple a cylinder to be filled with therapeutic gas;  
a gas sense device coupled to the fill port, the gas sense device operable to detect content  
of gas within the cylinder prior to filling; ~~The apparatus as defined in claim 31~~  
further comprising:  
a compressor coupled to the cylinder connector, the compressor operable to evacuate  
contents the cylinder; and  
wherein the compressor evacuates the cylinder if the content of the gas within the  
cylinder detected by the gas sense device is not suitable for therapeutic use.
36. (Original) The apparatus as defined in claim 35 further comprising:  
an oxygen concentrator providing therapeutic gas to the intensifier; and  
wherein the compressor is part of the oxygen concentrator.
37. (Currently Amended) A system comprising:  
a valve having an inlet port fluidly coupled to a source of therapeutic gas at a first  
pressure, the valve having a first outlet and a second outlet, and wherein the

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

therapeutic gas may be selectively permitted to flow to only one of the first and second outlets;  
an adjustable flow control device fluidly coupled to the first outlet, the adjustable flow control device operable to create a continuous flow of therapeutic gas at a selected flow rate;  
an intensifier fluidly coupled to the second outlet, the intensifier operable to produce therapeutic gas at a second pressure higher than the first pressure;  
a conserver fluidly coupled to the therapeutic gas as-at the second pressure, the conserver operable to release a bolus of therapeutic gas during a patient inhalation;  
a patient outlet port fluidly coupled to the adjustable flow control device, and the patient outlet port also fluidly coupled to the conserver; and  
wherein the volume of therapeutic gas released by the conserver is controlled by the selected flow rate of the adjustable flow device.

38. (Original) The system as defined in claim 37 further comprising:  
a controller electrically coupled to the adjustable flow control device and the conserver;  
and  
wherein the controller is operable to sense a flow rate setting of the adjustable flow control device, and wherein the controller is further operable to set volume of therapeutic gas released by the conserver based on the flow rate setting of the adjustable flow control device.
39. (New) A method comprising:  
compressing a stream of low-pressure therapeutic gas provided by an oxygen concentration system or other source of therapeutic gas, the compressing forms a compressed therapeutic gas stream;  
providing a first portion of the compressed therapeutic gas stream to fill a cylinder; and  
providing a second portion of the compressed therapeutic gas stream to a patient as a bolus of therapeutic gas.



10/04/2004 14:38 FAX 7132388008

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

40. (New) The method as defined in claim 39 further comprising setting a volume of the bolus of therapeutic gas based on a sensed setting for a continuous flow of therapeutic gas through a flow meter.
41. (New) A trans-fill system comprising:  
a compressor operable to increase pressure of therapeutic gas to create a compressed therapeutic gas stream having sufficient pressure to fill a cylinder;  
a conserver coupled to the compressed therapeutic gas stream, the conserver operable to deliver a bolus of therapeutic gas during inhalation of a patient; and  
a cylinder connector operable to couple a portable cylinder to the compressed therapeutic gas stream;  
wherein the trans-fill system is operable to provide therapeutic gas to the cylinder connector to fill the portable cylinder while providing therapeutic gas to the patient through the conserver.
42. (New) The trans-fill system as defined in claim 41 wherein the pressure of the compressed therapeutic gas stream is above 40 PSI.
43. (New) The trans-fill system as defined in claim 42 wherein the pressure of the compressed therapeutic gas stream is approximately 2200 PSI.
44. (New) A trans-fill system comprising:  
an oxygen concentrator comprising a first compressor that creates a therapeutic gas;  
a second compressor fluidly coupled to the oxygen concentrator that increases pressure of therapeutic gas to create a compressed therapeutic gas stream;  
a conserver coupled to the compressed therapeutic gas stream, the conserver operable to deliver a bolus of therapeutic gas during inhalation of a patient; and  
a cylinder connector operable to couple a portable cylinder to the compressed therapeutic gas stream;

10/04/2004 14:38 FAX 7132388008

Appl. No. 10/636,055  
Amdt. Dated October 4, 2004  
Reply to Office Action of September 21, 2004

wherein the trans-fill system is operable to provide therapeutic gas to the cylinder connector to fill the portable cylinder while providing therapeutic gas to the patient through the conserver.

45. (New) The trans-fill system as defined in claim 44 further comprising:  
a gas sense device fluidly coupled to the cylinder connector, the gas sense device operable to detect purity of therapeutic gas; and  
wherein the trans-fill system is operable to allow a portion of the gas within a connected portable cylinder to flow to the gas sense device, and wherein the trans-fill system refrains from filling the connected portable cylinder if the purity of the gas in the bottle, as determined by the gas sense device, falls below a predetermined threshold.

**This Page is Inserted by IFW Indexing and Scanning  
Operations and is not part of the Official Record**

## **BEST AVAILABLE IMAGES**

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

- ☐ **BLACK BORDERS**
- ☐ **IMAGE CUT OFF AT TOP, BOTTOM OR SIDES**
- ☐ **FADED TEXT OR DRAWING**
- ☐ **BLURRED OR ILLEGIBLE TEXT OR DRAWING**
- ☐ **SKEWED/SLANTED IMAGES**
- ☐ **COLOR OR BLACK AND WHITE PHOTOGRAPHS**
- ☐ **GRAY SCALE DOCUMENTS**
- ☐ **LINES OR MARKS ON ORIGINAL DOCUMENT**
- ☐ **REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY**
- ☐ **OTHER:** \_\_\_\_\_

**IMAGES ARE BEST AVAILABLE COPY.**

**As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.**